

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3935

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2024 10:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.5 F 0.0  
2:00 151.7 F 1.2  
4:00 155.7 F 4.1  
6:00 179.0 F 7.8  
8:00 197.7 F 12.0  
10:00 218.9 F 5.9  
12:00 231.7 F 10.3  
14:00 232.7 F 8.9  
16:00 240.2 F 11.9  
18:00 238.4 F 10.2  
20:00 250.8 F 16.2  
22:00 261.3 F 22.2  
24:00 271.0 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 271.8 F 29.9  
2:00 272.0 F 29.6  
3:00 272.3 F 30.2  
4:00 271.5 F 29.5

Min 270.6 F 28.9  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 235.7 F 0.0  
10:00 225.2 F 0.0  
15:00 228.0 F 0.0  
20:00 226.5 F 0.0  
25:00 227.2 F 0.0  
30:00 221.7 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 24:43  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:03:27  
CYCLE COMPLETE

09/12/2024 11:37 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_