
Midmark M11 - v1.0.0
Total Cycles: 3891

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2023 01:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.8 F 0.0
2:00 179.3 F 3.7
4:00 199.9 F 8.4
6:00 221.2 F 8.8
8:00 229.5 F 9.1
10:00 237.7 F 9.7
12:00 243.2 F 12.5
14:00 241.8 F 11.3
16:00 255.3 F 18.2
18:00 266.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.9
2:00 271.5 F 28.7
3:00 272.3 F 28.9
4:00 271.4 F 28.5

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 219.9 F 0.0
10:00 233.9 F 0.0
15:00 228.5 F 0.0
20:00 226.0 F 0.0
25:00 225.1 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

10/11/2023 02:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____