
Midmark M11 - v1.0.5
Total Cycles: 3565

Name: RITTER2
Sterilizer ID: V2313228
04/26/2024 01:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.2 F 0.0
2:00 112.6 F 0.9
4:00 127.4 F 3.0
6:00 152.8 F 5.8
8:00 176.9 F 9.2
10:00 197.4 F 12.9
12:00 224.4 F 6.0
14:00 232.8 F 10.2
16:00 231.0 F 8.4
18:00 237.2 F 11.5
20:00 236.9 F 9.5
22:00 249.5 F 15.6
24:00 260.4 F 21.7
26:00 270.7 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 273.4 F 29.8
2:00 271.7 F 29.6
3:00 272.7 F 29.9
4:00 271.2 F 29.3

Min 270.9 F 28.9
Max 273.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.2 F 0.0
5:00 237.0 F 0.0
10:00 225.9 F 0.0
15:00 178.8 F 0.0
20:00 174.9 F 0.0
25:00 174.1 F 0.0
30:00 173.4 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 26:42
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:05:27
CYCLE COMPLETE

04/26/2024 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____