
Midmark M11 - v1.0.0
Total Cycles: 1054

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2021 12:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 185.8 F 3.4
4:00 203.1 F 8.2
6:00 219.7 F 9.8
8:00 231.2 F 9.3
10:00 237.3 F 9.1
12:00 245.2 F 13.5
14:00 244.6 F 12.7
16:00 257.9 F 19.7
18:00 269.6 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.2 F 28.5
2:00 272.7 F 29.0
3:00 272.1 F 28.8
4:00 272.3 F 28.6

Min 271.7 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 231.5 F 0.0
10:00 230.0 F 0.0
15:00 228.6 F 0.0
20:00 227.7 F 0.0
25:00 196.6 F 0.0
30:00 188.0 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

07/26/2021 01:20 PM
ACCEPTED: X
REJECTED:
Operator: U01