

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005505  
10:45 10/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:56  
PRESSURIZING 2:00  
STERILIZING 2:57  
136.4 C 223kPa 2:57  
136.3 C 223kPa 3:27  
136.3 C 219kPa 3:57  
136.2 C 223kPa 4:27  
136.1 C 223kPa 4:57  
136.4 C 224kPa 5:27  
136.4 C 223kPa 5:57  
136.4 C 222kPa 6:27  
136.4 C 222kPa 6:27  
VENTING 6:27  
AIR DRYING 7:19  
CYCLE COMPLETE 20:05

-----  
Digital Signature #  
8AB04F0EE3D11361  
-----