

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3032

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2023 11:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.5 F 0.0  
2:00 137.1 F 2.3  
4:00 156.8 F 5.2  
6:00 183.3 F 9.1  
8:00 202.6 F 10.1  
10:00 221.7 F 6.7  
12:00 233.4 F 11.3  
14:00 232.9 F 7.6  
16:00 243.9 F 13.2  
18:00 241.8 F 11.2  
20:00 254.0 F 17.5  
22:00 265.9 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.4 F 28.6  
2:00 272.6 F 29.1  
3:00 271.3 F 28.3  
4:00 272.2 F 29.0

Min 271.3 F 28.3  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.8 F 0.0  
5:00 241.4 F 0.0  
10:00 234.9 F 0.0  
15:00 234.0 F 0.0  
20:00 229.3 F 0.0  
25:00 228.9 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 23:36  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:01:41  
CYCLE COMPLETE

03/24/2023 12:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_