

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000806  
14:26 18/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:57  
PRESSURIZING 7:13  
STERILIZING 9:10  
136.5 C 223kPa 9:10  
136.3 C 223kPa 9:40  
136.4 C 223kPa 10:10  
136.1 C 219kPa 10:40  
136.2 C 222kPa 11:10  
136.4 C 220kPa 11:40  
136.2 C 222kPa 12:10  
136.4 C 224kPa 12:40  
136.4 C 224kPa 12:40  
VENTING 12:41  
AIR DRYING 14:03  
CYCLE COMPLETE 24:54

-----  
Digital Signature #  
76CF87E49669DD35  
-----