
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007191
6:24 12/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:05
STERILIZING 4:02
136.4 C 220kPa 4:02
136.4 C 220kPa 4:32
136.4 C 222kPa 5:02
136.4 C 220kPa 5:32
136.4 C 221kPa 6:02
136.2 C 219kPa 6:32
136.4 C 221kPa 7:02
135.9 C 219kPa 7:32
135.9 C 219kPa 7:32
VENTING 7:32
AIR DRYING 8:20
CASSETTE REMOVED
CYCLE COMPLETE 33:02

Digital Signature #
57E9BB766B7B8FAA
