
Midmark M11 - v1.0.0
Total Cycles: 4355

Name: RITTER 1
Sterilizer ID: V2178198
03/04/2024 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 68.9 F 0.2
4:00 77.4 F 1.0
6:00 100.2 F 3.1
8:00 134.3 F 5.7
10:00 162.1 F 8.9
12:00 183.9 F 12.3
14:00 202.3 F 2.9
16:00 211.1 F 5.9
18:00 221.6 F 9.1
20:00 230.8 F 12.8
22:00 231.8 F 8.0
24:00 243.3 F 13.1
26:00 241.3 F 11.0
28:00 253.2 F 16.9
30:00 263.5 F 23.0
32:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 28.9
2:00 271.4 F 28.4
3:00 272.5 F 28.8
4:00 271.3 F 28.3

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 216.4 F 0.0
10:00 223.9 F 0.0
15:00 233.9 F 0.0
20:00 227.2 F 0.0
25:00 227.2 F 0.0
30:00 226.7 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:15
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:33
CYCLE COMPLETE

03/04/2024 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____