
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001397
11:40 29/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:40
STERILIZING 4:54
136.0 C 221kPa 4:54
136.4 C 225kPa 5:25
136.3 C 223kPa 5:55
135.9 C 221kPa 6:25
136.4 C 224kPa 6:55
136.4 C 224kPa 7:25
136.4 C 224kPa 7:55
136.4 C 224kPa 8:25
136.4 C 224kPa 8:25
VENTING 8:25
AIR DRYING 9:20
CYCLE COMPLETE 24:20

Digital Signature #
50CD10724B8717B3
