
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002780
17:32 01/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:37
STERILIZING 3:33
136.0 C 220kPa 3:33
136.4 C 225kPa 4:03
136.2 C 222kPa 4:33
136.4 C 221kPa 5:03
136.2 C 223kPa 5:33
136.4 C 223kPa 6:03
136.3 C 223kPa 6:33
136.4 C 222kPa 7:03
136.4 C 222kPa 7:03
VENTING 7:03
AIR DRYING 7:57
CYCLE COMPLETE 8:26

Digital Signature #
D3FAD62BB268A0BE
