
Midmark M11 - v1.0.5
Total Cycles: 1244

Name: RITTER2
Sterilizer ID: V2313228
06/29/2022 03:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.8 F 0.0
2:00 191.3 F 4.1
4:00 207.0 F 9.3
6:00 225.8 F 7.7
8:00 234.2 F 10.8
10:00 232.4 F 8.1
12:00 241.1 F 10.9
14:00 246.8 F 13.7
16:00 258.8 F 20.4
18:00 269.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.7 F 29.3
2:00 272.4 F 29.8
3:00 271.9 F 29.0
4:00 272.6 F 29.7

Min 271.5 F 29.0
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 231.0 F 0.0
10:00 232.3 F 0.0
15:00 231.4 F 0.0
20:00 226.2 F 0.0
25:00 208.9 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

06/29/2022 04:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____