
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000961
12:16 07/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:50
STERILIZING 5:07
136.0 C 222kPa 5:07
136.2 C 232kPa 5:37
136.4 C 233kPa 6:07
136.3 C 232kPa 6:37
136.3 C 232kPa 7:07
136.3 C 231kPa 7:37
136.3 C 232kPa 8:07
136.3 C 232kPa 8:37
136.3 C 232kPa 8:37
VENTING 8:37
AIR DRYING 9:34
CYCLE COMPLETE 24:34

Digital Signature #
59454E1E3F186B92
