

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001678  
11:33 10/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:17  
PRESSURIZING 2:24  
STERILIZING 3:25  
136.0 C 220kPa 3:25  
136.0 C 222kPa 3:55  
136.3 C 224kPa 4:25  
136.4 C 225kPa 4:55  
136.3 C 221kPa 5:25  
136.5 C 223kPa 5:55  
136.2 C 224kPa 6:25  
136.2 C 223kPa 6:55  
136.2 C 223kPa 6:55  
VENTING 6:55  
AIR DRYING 7:44  
CYCLE COMPLETE 7:46  
-----

Digital Signature #  
65C94DEBE983692B  
-----