

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007862  
15:15 11/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:51  
PRESSURIZING 2:56  
STERILIZING 3:57  
136.0 C 220kPa 3:57  
136.0 C 219kPa 4:27  
136.3 C 222kPa 4:57  
135.9 C 221kPa 5:27  
136.1 C 220kPa 5:57  
135.8 C 220kPa 6:27  
136.4 C 224kPa 6:57  
136.4 C 222kPa 7:27  
136.4 C 222kPa 7:27  
VENTING 7:27  
AIR DRYING 8:17  
CYCLE COMPLETE 13:44  
-----

Digital Signature #  
C2C3FCEEFB4F5B35  
-----