
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001056
14:16 04/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:25
PRESSURIZING 4:26
STERILIZING 5:44
136.5 C 224kPa 5:44
136.3 C 224kPa 6:14
136.4 C 225kPa 6:44
136.4 C 223kPa 7:14
136.3 C 225kPa 7:44
136.2 C 224kPa 8:14
136.4 C 224kPa 8:44
136.3 C 224kPa 9:14
136.3 C 224kPa 9:14
VENTING 9:14
AIR DRYING 10:11
CYCLE COMPLETE 25:11

Digital Signature #
073189588F295127
