

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3003

Name: RITTER 1  
Sterilizer ID: V2178198  
03/14/2023 11:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
2:00 183.2 F 3.7  
4:00 201.9 F 8.1  
6:00 219.2 F 12.3  
8:00 228.4 F 9.0  
10:00 238.1 F 9.6  
12:00 242.7 F 12.6  
14:00 242.2 F 11.5  
16:00 255.6 F 18.4  
18:00 267.2 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.2  
2:00 271.4 F 28.4  
3:00 272.7 F 29.4  
4:00 272.1 F 28.8  
  
Min 271.3 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.3 F 0.0  
5:00 216.3 F 0.0  
10:00 228.4 F 0.0  
15:00 244.6 F 0.0  
20:00 222.7 F 0.0  
25:00 212.2 F 0.0  
30:00 209.3 F 0.0  
  
Drying Complete

FILLING: 1:28  
HEATING: 19:25  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:07  
CYCLE COMPLETE

03/14/2023 12:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_