
Midmark M11 - v1.0.0
Total Cycles: 2899

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2023 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.8 F 0.0
2:00 151.3 F 3.1
4:00 174.5 F 6.8
6:00 196.7 F 11.1
8:00 211.0 F 4.4
10:00 224.3 F 9.3
12:00 233.5 F 10.4
14:00 237.4 F 10.8
16:00 234.0 F 8.2
18:00 248.7 F 14.9
20:00 260.0 F 21.3
22:00 270.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.5
2:00 272.0 F 28.8
3:00 271.6 F 28.9
4:00 272.6 F 29.4

Min 271.4 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 210.4 F 0.0
10:00 232.7 F 0.0
15:00 230.4 F 0.0
20:00 228.8 F 0.0
25:00 226.3 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:50
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:11
CYCLE COMPLETE

02/14/2023 02:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____