

-----  
Midmark M11 - v1.0.5  
Total Cycles: 951

Name: RITTER2  
Sterilizer ID: V2313228  
04/04/2022 11:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.8 F 0.0  
2:00 174.0 F 0.3  
4:00 190.3 F 3.3  
6:00 213.0 F 8.1  
8:00 227.5 F 12.5  
10:00 234.8 F 10.0  
12:00 236.3 F 8.8  
14:00 242.0 F 11.2  
16:00 247.8 F 14.5  
18:00 260.7 F 21.7  
20:00 272.2 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.7 F 29.5  
2:00 272.0 F 28.8  
3:00 271.4 F 28.9  
4:00 272.6 F 29.5

Min 271.2 F 28.5  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.8 F 0.0  
5:00 231.8 F 0.0  
10:00 221.6 F 0.0  
15:00 222.3 F 0.0  
20:00 218.6 F 0.0  
25:00 212.3 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 20:32  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:43  
CYCLE COMPLETE

04/04/2022 12:42 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02