

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3991

Name: RITTER 1  
Sterilizer ID: V2178198  
11/07/2023 02:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.3 F 0.0  
2:00 139.7 F 2.1  
4:00 157.5 F 5.3  
6:00 181.6 F 9.0  
8:00 201.4 F 13.2  
10:00 216.8 F 5.6  
12:00 227.3 F 9.8  
14:00 232.9 F 9.2  
16:00 239.4 F 11.1  
18:00 236.3 F 9.0  
20:00 249.9 F 15.3  
22:00 261.1 F 21.6  
24:00 271.6 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.6 F 29.1  
2:00 271.8 F 28.4  
3:00 271.6 F 28.7  
4:00 272.5 F 29.0

Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 209.8 F 0.0  
10:00 215.6 F 0.0  
15:00 222.4 F 0.0  
20:00 224.4 F 0.0  
25:00 225.9 F 0.0  
30:00 227.2 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 24:39  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:03:26  
CYCLE COMPLETE

11/07/2023 03:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_