
Midmark M11 - v1.0.0
Total Cycles: 5258

Name: RITTER 1
Sterilizer ID: V2178198
01/22/2025 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.5 F 0.0
2:00 140.7 F 1.6
4:00 154.6 F 4.8
6:00 179.9 F 8.3
8:00 198.6 F 12.3
10:00 213.8 F 3.8
12:00 222.8 F 7.8
14:00 231.8 F 11.9
16:00 234.0 F 7.8
18:00 245.0 F 13.0
20:00 242.0 F 10.9
22:00 253.6 F 16.7
24:00 264.0 F 22.8
26:00 272.0 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 271.9 F 28.7
2:00 272.2 F 28.4
3:00 272.4 F 28.9
4:00 271.9 F 28.3

Min 270.7 F 27.8
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 214.0 F 0.0
10:00 192.8 F 0.0
15:00 190.7 F 0.0
20:00 189.8 F 0.0
25:00 189.3 F 0.0
30:00 194.2 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 26:06
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:03
CYCLE COMPLETE

01/22/2025 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____