

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1458

Name: RITTER2  
Sterilizer ID: V2313228  
09/06/2022 11:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.5 F 0.0  
2:00 180.9 F 2.6  
4:00 188.5 F 7.4  
6:00 208.7 F 13.0  
8:00 226.2 F 7.8  
10:00 239.0 F 13.2  
12:00 239.4 F 11.0  
14:00 237.7 F 9.7  
16:00 252.1 F 16.7  
18:00 263.8 F 23.6  
20:00 271.9 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.9 F 29.7  
2:00 272.6 F 29.1  
3:00 271.5 F 28.9  
4:00 272.8 F 29.7

Min 271.5 F 28.7  
Max 273.1 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
5:00 231.1 F 0.0  
10:00 219.4 F 0.0  
15:00 222.6 F 0.0  
20:00 206.6 F 0.0  
25:00 204.7 F 0.0  
30:00 202.4 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 20:07  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

09/06/2022 12:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_