
Midmark M11 - v1.0.0
Total Cycles: 2852

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2023 04:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.0 F 0.0
2:00 162.8 F 3.2
4:00 184.1 F 7.2
6:00 210.4 F 11.6
8:00 214.2 F 5.4
10:00 226.8 F 10.3
12:00 230.5 F 7.5
14:00 241.1 F 12.2
16:00 238.8 F 10.4
18:00 252.3 F 16.9
20:00 264.0 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.3 F 29.3
2:00 271.2 F 28.7
3:00 272.2 F 29.5
4:00 271.7 F 28.9

Min 271.1 F 28.6
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 209.4 F 0.0
10:00 220.7 F 0.0
15:00 228.9 F 0.0
20:00 226.6 F 0.0
25:00 229.7 F 0.0
30:00 229.9 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:23
CYCLE COMPLETE

02/02/2023 05:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____