

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3456

Name: RITTER2  
Sterilizer ID: V2313228  
03/15/2024 10:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.0 F 0.0  
2:00 141.9 F 0.5  
4:00 143.5 F 2.4  
6:00 171.4 F 6.2  
8:00 197.3 F 10.4  
10:00 214.7 F 6.1  
12:00 224.6 F 9.3  
14:00 233.8 F 10.6  
16:00 237.6 F 11.3  
18:00 237.3 F 9.8  
20:00 250.2 F 16.2  
22:00 261.6 F 22.7  
24:00 272.1 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.8  
1:00 271.9 F 29.7  
2:00 272.2 F 29.7  
3:00 272.5 F 29.9  
4:00 272.7 F 30.2

Min 271.6 F 29.7  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 211.8 F 0.0  
10:00 217.1 F 0.0  
15:00 222.2 F 0.0  
20:00 218.6 F 0.0  
25:00 215.3 F 0.0  
30:00 215.8 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:13  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:03:44  
CYCLE COMPLETE

03/15/2024 11:25 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_