

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1599

Name: RITTER2  
Sterilizer ID: V2313228  
10/14/2022 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.3 F 0.0  
2:00 167.5 F 3.5  
4:00 179.5 F 7.7  
6:00 202.6 F 13.1  
8:00 220.2 F 7.2  
10:00 234.3 F 12.7  
12:00 236.7 F 10.2  
14:00 235.8 F 8.5  
16:00 250.0 F 15.6  
18:00 261.9 F 22.4  
20:00 272.8 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.1  
2:00 272.1 F 29.4  
3:00 272.0 F 28.9  
4:00 272.2 F 29.5

Min 271.5 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
5:00 215.7 F 0.0  
10:00 230.1 F 0.0  
15:00 233.9 F 0.0  
20:00 225.2 F 0.0  
25:00 220.5 F 0.0  
30:00 219.6 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

10/14/2022 12:36 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_