
Midmark M11 - v1.0.5
Total Cycles: 3655

Name: RITTER2
Sterilizer ID: V2313228
05/29/2024 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 76.1 F 0.0
4:00 82.5 F 0.1
6:00 109.9 F 1.7
8:00 138.2 F 4.5
10:00 165.8 F 8.0
12:00 188.0 F 11.4
14:00 206.2 F 3.7
16:00 215.4 F 7.2
18:00 225.8 F 11.1
20:00 228.1 F 6.9
22:00 239.5 F 12.0
24:00 237.8 F 10.0
26:00 250.4 F 16.1
28:00 261.3 F 22.4
30:00 271.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.6 F 29.7
2:00 271.3 F 29.1
3:00 272.6 F 29.8
4:00 271.5 F 29.0

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 208.6 F 0.0
10:00 184.8 F 0.0
15:00 182.9 F 0.0
20:00 178.5 F 0.0
25:00 176.3 F 0.0
30:00 175.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:38
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

05/29/2024 10:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____