
Midmark M11 - v1.0.5
Total Cycles: 4095

Name: RITTER2
Sterilizer ID: V2313228
11/07/2024 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.4 F 0.0
2:00 72.9 F 0.0
4:00 81.2 F 0.0
6:00 111.0 F 0.0
8:00 139.8 F 0.1
10:00 159.3 F 0.2
12:00 181.8 F 0.6
14:00 201.2 F 3.5
16:00 219.5 F 6.5
18:00 223.9 F 9.9
20:00 232.5 F 11.4
22:00 233.5 F 9.6
24:00 240.0 F 11.0
26:00 242.7 F 12.3
28:00 241.1 F 11.1
30:00 253.0 F 17.1
32:00 263.3 F 23.2
34:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 271.9 F 29.2
2:00 272.4 F 29.2
3:00 271.5 F 28.6
4:00 272.4 F 29.5

Min 271.4 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 227.6 F 0.0
10:00 218.9 F 0.0
15:00 212.2 F 0.0
20:00 212.4 F 0.0
25:00 220.1 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:16
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:12:49
CYCLE COMPLETE

11/07/2024 11:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____