
Midmark M11 - v1.0.0
Total Cycles: 3723

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2023 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.5 F 0.0
2:00 178.0 F 1.2
4:00 186.1 F 4.4
6:00 206.8 F 9.1
8:00 221.9 F 8.4
10:00 230.3 F 9.1
12:00 237.8 F 10.5
14:00 241.6 F 11.8
16:00 239.6 F 10.2
18:00 252.8 F 16.8
20:00 264.2 F 23.4
22:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.0
2:00 271.4 F 28.3
3:00 272.6 F 29.1
4:00 271.8 F 28.4

Min 271.4 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 218.5 F 0.0
10:00 220.6 F 0.0
15:00 219.3 F 0.0
20:00 220.1 F 0.0
25:00 220.3 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

09/06/2023 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____