
Midmark M11 - v1.0.0
Total Cycles: 5487

Name: RITTER 1
Sterilizer ID: V2178198
04/08/2025 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 88.4 F 0.0
2:00 93.0 F 0.5
4:00 125.3 F 2.1
6:00 144.7 F 4.8
8:00 153.6 F 7.9
10:00 173.7 F 11.3
12:00 206.5 F 2.8
14:00 209.5 F 4.6
16:00 220.2 F 7.8
18:00 230.4 F 11.5
20:00 231.2 F 7.1
22:00 241.2 F 12.2
24:00 239.0 F 10.0
26:00 250.5 F 15.6
28:00 260.5 F 21.3
30:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 271.8 F 29.1
2:00 272.5 F 29.4
3:00 272.3 F 29.1
4:00 272.5 F 29.3

Min 271.0 F 28.4
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 205.7 F 0.0
10:00 198.9 F 0.0
15:00 207.9 F 0.0
20:00 214.6 F 0.0
25:00 209.6 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:06
STERILIZING: 4:00
VENTING: 4:02
DRYING: 30:00
TOTAL CYCLE: 01:09:55
CYCLE COMPLETE

04/08/2025 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____