
Midmark M11 - v1.0.5
Total Cycles: 1493

Name: RITTER2
Sterilizer ID: V2313228
09/15/2022 06:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 79.8 F 0.0
4:00 118.1 F 0.0
6:00 128.2 F 0.1
8:00 141.6 F 1.9
10:00 165.1 F 4.9
12:00 184.7 F 8.1
14:00 211.1 F 11.8
16:00 221.8 F 5.5
18:00 230.8 F 9.7
20:00 235.8 F 10.1
22:00 241.0 F 11.3
24:00 238.1 F 9.6
26:00 251.2 F 15.9
28:00 262.3 F 22.4
30:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.6
1:00 272.3 F 29.1
2:00 272.0 F 28.9
3:00 272.9 F 29.5
4:00 271.8 F 29.1

Min 271.0 F 28.4
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 188.5 F 0.0
10:00 178.0 F 0.0
15:00 164.1 F 0.0
20:00 163.6 F 0.0
25:00 161.3 F 0.0
30:00 158.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 30:26
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:39
CYCLE COMPLETE

09/15/2022 07:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____