
Midmark M11 - v1.0.5
Total Cycles: 408

Name: RITTER2
Sterilizer ID: V2313228
09/30/2021 06:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 73.5 F 0.0
4:00 81.1 F 0.0
6:00 111.0 F 0.2
8:00 155.5 F 0.6
10:00 179.2 F 3.9
12:00 207.9 F 7.6
14:00 219.0 F 11.5
16:00 223.9 F 5.2
18:00 230.7 F 9.7
20:00 239.4 F 10.4
22:00 241.4 F 11.7
24:00 239.2 F 10.2
26:00 252.1 F 16.6
28:00 262.9 F 22.9
30:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.8
1:00 272.6 F 29.6
2:00 271.5 F 28.9
3:00 272.6 F 29.3
4:00 271.7 F 29.1

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 234.9 F 0.0
10:00 221.4 F 0.0
15:00 228.4 F 0.0
20:00 209.1 F 0.0
25:00 208.8 F 0.0
30:00 205.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 30:15
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:08:21
CYCLE COMPLETE

09/30/2021 08:05 AM
ACCEPTED: X
REJECTED:
Operator: U02