

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000758  
13:27 27/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:04  
STERILIZING 4:12  
136.4 C 233kPa 4:12  
136.1 C 231kPa 4:43  
136.4 C 232kPa 5:13  
136.4 C 233kPa 5:43  
136.4 C 231kPa 6:13  
136.3 C 232kPa 6:43  
136.3 C 233kPa 7:13  
136.3 C 232kPa 7:43  
136.3 C 232kPa 7:43  
VENTING 7:43  
AIR DRYING 8:42  
CYCLE COMPLETE 8:44  
-----

Digital Signature #  
F38A289CA053555F  
-----