
Midmark M11 - v1.0.5
Total Cycles: 1878

Name: RITTER2
Sterilizer ID: V2313228
12/22/2022 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	136.6 F	0.0
2:00	137.7 F	0.0
4:00	150.5 F	0.0
6:00	187.4 F	0.7
8:00	210.6 F	5.4
10:00	223.3 F	10.2
12:00	230.1 F	7.7
14:00	238.8 F	11.9
16:00	237.6 F	9.9
18:00	237.6 F	8.7
20:00	251.2 F	16.0
22:00	263.2 F	23.1
24:00	271.9 F	28.5

STERILIZING

mm:ss	Degrees	PSI
0:00	271.5 F	28.7
1:00	272.4 F	29.1
2:00	271.5 F	28.9
3:00	272.4 F	29.1
4:00	271.5 F	28.6

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

mm:ss	Drying Start	Degrees	PSI
0:00		192.9 F	0.0
5:00		210.9 F	0.0
10:00		227.3 F	0.0
15:00		230.0 F	0.0
20:00		229.7 F	0.0
25:00		226.2 F	0.0
30:00		222.7 F	0.0

Drying Complete

FILLING: 0:60
HEATING: 24:07
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:14
CYCLE COMPLETE

12/22/2022 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____