
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001282
14:01 03/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 4:30
STERILIZING 5:47
136.0 C 221kPa 5:47
136.2 C 223kPa 6:17
136.2 C 223kPa 6:47
136.4 C 224kPa 7:17
136.3 C 223kPa 7:47
136.3 C 224kPa 8:17
136.4 C 224kPa 8:47
136.2 C 223kPa 9:17
136.2 C 223kPa 9:17
VENTING 9:17
AIR DRYING 10:15
CYCLE COMPLETE 12:48

Digital Signature #
1746F6680A00A13E
