
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002229
11:45 06/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 2:00
STERILIZING 3:04
136.5 C 224kPa 3:04
136.4 C 226kPa 3:34
136.3 C 225kPa 4:04
136.2 C 225kPa 4:34
136.4 C 225kPa 5:04
136.4 C 226kPa 5:34
136.2 C 224kPa 6:04
136.4 C 226kPa 6:34
136.4 C 226kPa 6:34
VENTING 6:34
AIR DRYING 7:26
CYCLE COMPLETE 37:26

Digital Signature #
E77196F553222C8C
