
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 004747
10:20 22/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:12
STERILIZING 3:08
136.2 C 220kPa 3:08
136.3 C 223kPa 3:38
136.4 C 223kPa 4:08
136.4 C 224kPa 4:38
136.4 C 222kPa 5:08
136.4 C 223kPa 5:38
136.3 C 221kPa 6:08
136.4 C 222kPa 6:38
136.4 C 222kPa 6:38
VENTING 6:38
AIR DRYING 7:26
CYCLE COMPLETE 7:35

Digital Signature #
617D4D7F81D4A907
