
Midmark M11 - v1.0.5
Total Cycles: 340

Name: RITTER2
Sterilizer ID: V2313228
09/08/2021 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.2 F 0.0
2:00 194.3 F 4.6
4:00 216.3 F 10.5
6:00 224.8 F 6.9
8:00 239.2 F 13.4
10:00 241.8 F 12.5
12:00 242.9 F 12.4
14:00 256.7 F 19.7
16:00 268.5 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.3 F 29.4
2:00 272.7 F 29.9
3:00 271.7 F 29.4
4:00 272.3 F 29.5

Min 271.6 F 29.1
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 217.8 F 0.0
10:00 210.1 F 0.0
15:00 209.2 F 0.0
20:00 212.9 F 0.0
25:00 191.3 F 0.0
30:00 201.8 F 0.0

Drying Complete

FILLING: 3:08
HEATING: 17:06
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:24
CYCLE COMPLETE

09/08/2021 12:08 PM
ACCEPTED: X
REJECTED:
Operator: U01