

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4938

Name: RITTER 1  
Sterilizer ID: V2178198  
09/25/2024 12:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.4 F 0.0  
2:00 120.4 F 0.0  
4:00 129.6 F 0.1  
6:00 169.4 F 0.2  
8:00 198.3 F 0.5  
10:00 210.9 F 3.5  
12:00 219.7 F 6.9  
14:00 230.0 F 10.6  
16:00 232.5 F 8.1  
18:00 241.2 F 11.8  
20:00 238.2 F 9.6  
22:00 238.0 F 9.1  
24:00 249.4 F 14.7  
26:00 261.2 F 21.2  
28:00 271.8 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.5 F 28.7  
2:00 271.9 F 28.1  
3:00 271.6 F 28.1  
4:00 272.6 F 28.8

Min 271.4 F 27.8  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
5:00 213.3 F 0.0  
10:00 192.3 F 0.0  
15:00 191.4 F 0.0  
20:00 208.1 F 0.0  
25:00 228.1 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 28:38  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:06:58  
CYCLE COMPLETE

09/25/2024 01:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_