

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3079

Name: RITTER2  
Sterilizer ID: V2313228  
11/02/2023 03:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.0 F 0.0  
2:00 166.8 F 3.2  
4:00 186.5 F 7.8  
6:00 209.4 F 13.4  
8:00 223.8 F 7.5  
10:00 241.8 F 13.2  
12:00 240.1 F 11.2  
14:00 239.8 F 10.4  
16:00 254.5 F 17.8  
18:00 267.0 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.8 F 29.1  
2:00 271.5 F 28.6  
3:00 272.9 F 29.2  
4:00 271.8 F 28.5  
  
Min 271.4 F 28.3  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 223.8 F 0.0  
10:00 209.2 F 0.0  
15:00 208.8 F 0.0  
20:00 207.9 F 0.0  
25:00 206.5 F 0.0  
30:00 203.7 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 19:21  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:49  
CYCLE COMPLETE

11/02/2023 04:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_