
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008068
11:52 14/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:06
STERILIZING 3:12
136.0 C 220kPa 3:12
136.4 C 224kPa 3:42
136.4 C 224kPa 4:12
136.2 C 218kPa 4:42
136.2 C 222kPa 5:12
136.3 C 224kPa 5:42
136.2 C 224kPa 6:12
136.2 C 222kPa 6:42
136.2 C 222kPa 6:42
VENTING 6:42
AIR DRYING 7:33
CYCLE COMPLETE 37:33

Digital Signature #
5B7366684F7141B6
