
Midmark M11 - v1.0.0
Total Cycles: 1448

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2021 09:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.6 F 0.0
2:00 140.9 F 0.0
4:00 158.0 F 0.1
6:00 188.8 F 2.7
8:00 201.0 F 7.1
10:00 217.1 F 11.7
12:00 224.4 F 6.7
14:00 235.9 F 11.8
16:00 235.3 F 9.1
18:00 238.7 F 9.5
20:00 247.0 F 13.8
22:00 259.2 F 20.4
24:00 270.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 28.9
2:00 271.3 F 28.4
3:00 272.6 F 29.0
4:00 271.5 F 28.3

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 215.6 F 0.0
10:00 238.1 F 0.0
15:00 237.5 F 0.0
20:00 233.5 F 0.0
25:00 223.8 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:55
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

12/13/2021 10:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____