
Midmark M11 - v1.0.0
Total Cycles: 2651

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2022 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.2 F 0.0
2:00 155.3 F 0.2
4:00 164.9 F 1.5
6:00 193.7 F 5.5
8:00 209.2 F 10.5
10:00 223.0 F 5.3
12:00 234.3 F 10.9
14:00 234.2 F 7.8
16:00 242.1 F 11.5
18:00 245.6 F 12.9
20:00 258.2 F 19.7
22:00 269.6 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.2 F 28.7
2:00 272.0 F 29.0
3:00 271.8 F 28.4
4:00 272.4 F 29.1

Min 271.5 F 28.3
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.8 F 0.0
5:00 212.7 F 0.0
10:00 240.1 F 0.0
15:00 238.2 F 0.0
20:00 222.9 F 0.0
25:00 218.0 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:58
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

12/07/2022 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____