
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001001
14:33 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 1:59
STERILIZING 3:05
136.4 C 233kPa 3:05
136.3 C 233kPa 3:35
136.3 C 232kPa 4:05
136.4 C 233kPa 4:35
136.4 C 233kPa 5:05
136.3 C 233kPa 5:35
136.4 C 233kPa 6:05
136.1 C 230kPa 6:35
136.1 C 230kPa 6:35
VENTING 6:35
AIR DRYING 7:26
CYCLE COMPLETE 9:44

Digital Signature #
E217A69EC2E674F6
