
Midmark M11 - v1.0.0
Total Cycles: 1470

Name: RITTER 1
Sterilizer ID: V2178198
12/16/2021 03:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.7 F 0.0
2:00 184.2 F 3.7
4:00 203.8 F 9.5
6:00 225.2 F 6.8
8:00 237.3 F 12.7
10:00 238.2 F 10.8
12:00 237.2 F 9.6
14:00 252.8 F 17.0
16:00 265.4 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.1 F 29.0
2:00 272.0 F 29.1
3:00 272.0 F 28.9
4:00 271.9 F 29.1

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 231.2 F 0.0
10:00 229.5 F 0.0
15:00 230.4 F 0.0
20:00 229.9 F 0.0
Dry Cycle Aborted

FILLING: 2:24
HEATING: 17:38
STERILIZING: 4:00
VENTING: 3:02
DRYING: 20:12
TOTAL CYCLE: 00:47:16

12/16/2021 04:04 PM
ACCEPTED: X
REJECTED:
Operator: U02