
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001527
11:14 21/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:00
136.0 C 220kPa 3:00
136.4 C 224kPa 3:30
136.4 C 225kPa 4:00
136.4 C 224kPa 4:30
136.4 C 225kPa 5:00
136.4 C 223kPa 5:30
136.4 C 225kPa 6:00
136.4 C 224kPa 6:30
136.4 C 224kPa 6:30
VENTING 6:30
AIR DRYING 7:21
CYCLE COMPLETE 7:24

Digital Signature #
0F4017F0B47DDA05
