

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000092  
13:23 13/02/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:24  
PRESSURIZING 1:31  
STERILIZING 2:51  
136.0 C 220kPa 2:51  
136.2 C 223kPa 3:21  
136.4 C 225kPa 3:51  
136.4 C 224kPa 4:21  
136.4 C 221kPa 4:51  
136.4 C 221kPa 5:21  
136.4 C 224kPa 5:51  
136.4 C 223kPa 6:21  
136.4 C 223kPa 6:21  
VENTING 6:21  
AIR DRYING 7:16  
CYCLE COMPLETE 7:21  
-----

Digital Signature #  
804A69806DB20B41  
-----