
Midmark M11 - v1.0.5
Total Cycles: 969

Name: RITTER2
Sterilizer ID: V2313228
04/07/2022 12:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.9 F 0.0
2:00 149.3 F 0.1
4:00 172.2 F 0.6
6:00 197.4 F 4.3
8:00 227.3 F 8.9
10:00 229.0 F 10.3
12:00 237.7 F 10.8
14:00 234.9 F 8.8
16:00 239.4 F 10.1
18:00 249.1 F 15.2
20:00 261.6 F 22.2
22:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.5 F 29.3
2:00 271.5 F 29.2
3:00 272.5 F 29.2
4:00 271.5 F 29.1

Min 271.2 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 214.4 F 0.0
10:00 193.0 F 0.0
15:00 193.4 F 0.0
20:00 193.6 F 0.0
25:00 202.1 F 0.0
30:00 211.3 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

04/07/2022 01:39 PM
ACCEPTED: X
REJECTED:
Operator: U02