
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005586
7:44 24/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:00
STERILIZING 4:00
136.4 C 220kPa 4:00
136.3 C 222kPa 4:30
136.4 C 222kPa 5:00
136.2 C 218kPa 5:30
136.2 C 220kPa 6:00
136.0 C 216kPa 6:30
136.3 C 223kPa 7:00
136.4 C 223kPa 7:30
136.4 C 223kPa 7:30
VENTING 7:30
AIR DRYING 8:21
CYCLE COMPLETE 10:37

Digital Signature #
435D899B9D1C877F
