
Midmark M11 - v1.0.0
Total Cycles: 2496

Name: RITTER 1
Sterilizer ID: V2178198
10/26/2022 04:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 179.0 F 4.3
4:00 194.9 F 8.8
6:00 215.6 F 9.1
8:00 224.7 F 9.1
10:00 234.5 F 9.5
12:00 240.5 F 12.0
14:00 240.0 F 10.7
16:00 253.5 F 17.5
18:00 265.5 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.4
2:00 271.5 F 28.8
3:00 272.6 F 29.5
4:00 271.8 F 28.8

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 243.8 F 0.0
10:00 231.5 F 0.0
15:00 230.9 F 0.0
20:00 229.2 F 0.0
25:00 229.9 F 0.0
30:00 231.4 F 0.0

Drying Complete

FILLING: 1:29
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

10/26/2022 05:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____