
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001014
15:23 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:14
STERILIZING 3:13
136.4 C 233kPa 3:13
136.4 C 233kPa 3:43
136.2 C 232kPa 4:13
136.1 C 229kPa 4:43
136.4 C 234kPa 5:13
136.3 C 233kPa 5:43
136.3 C 233kPa 6:13
136.4 C 233kPa 6:43
136.4 C 233kPa 6:43
VENTING 6:43
AIR DRYING 7:35
CYCLE COMPLETE 7:40

Digital Signature #
0A41F78D46C6892D
